

G900 Datasheet



G·S·L

WWW.GLOVEMAN.CO.UK



Gloves tested to comply with Medical Devices Directives 93/42/EEC and European Standards EN455 parts 1, 2 and 3 and EN374. Designed for medical procedures, offering a high quality and dependable latex free glove.

- Visually inspected for holes and minor defects
- Random samples are filled with 1000ml of water and examined for leaks
- Tested to AQL 1.5
- Manufactured from nitrile butadiene
- Mildly chlorinated and thoroughly washed
- Typical force at break: 9 Newtons

Recommended for use with Medical, Laboratory, Tattooists, Chemical handling, Veterinary and Emergency services .

Dimensions	Small	Medium	Large	Extra Large
Palm Width (mm)	85±10mm	95±10mm	105±10mm	>115mm±10mm
Length (mm)	Min 240mm	Min 240mm	Min 240mm	Min 240mm
Thickness (mm) palm	0.07±0.02	0.07±0.02	0.07±0.02	0.07±0.02
Thickness (mm) finger	0.10±0.02	0.10±0.02	0.10±0.02	0.10±0.02

Benefits	
Latex free	Reducing the risk of hypersensitivity and allergic reactions
Low in residual chemicals	Reduces the risk of contact dermatitis associated with these chemical
Flexible materials	For improved feel and fit reducing fatigue during extended wear
Textured	For improved wet grip
Powder free	Reduces potential dust contamination
Beaded cuff	For added strength and easier donning
Available in 4 sizes	Providing wearer with a range of comfortable sizes

Characteristics	Unaged	Aged
Force at break	19 Mpa	17 Mpa
Elongation at break	600%	550%

Black Powder Free Nitrile Exam Gloves

Gloveman Supplies Ltd

Vision House
Treleigh Ind Est
Redruth TR16 4AX
t. 01209 314759
f. 01209 314760

e. sales@gloveman.co.uk

www.gloveman.co.uk