

GN400 Datasheet



AQL 1.5

EN455 Parts 1, 2 & 3

MDD 93/42/EEC

EN374

Class 1 Medical Device



Stretch Nitrile Lilac Powder Free Gloves

General Details

Material	Stretch Nitrile
Surface Finish	Powder Free (inner chlorinated., outer polymer coated)
Colour	Lilac
Geometry	Ambidextrous
Texture	Fully Textured (Palm Textured)
Cuff and Finishing	Beaded

Glove Specifications

Size	Weight g/pc	Length mm	Palm Width mm	Cuff	Thickness Palm	Finger
XS	2.7	240	≤ 80	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
S	2.8	240	85 ± 5	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
M	3.2	240	95 ± 5	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
L	3.7	240	105 ± 5	0.095 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
XL	4.2	240	≥ 110	0.05 ± 0.01	0.08 ± 0.01	0.09 ± 0.01
Powder Content (Maximum)			2 mg / glove			
Protein Content (Maximum)			Not applicable			

Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

Manufacturing Standards

Manufacturing Standards	Incompliance with ISO 9001:2008
Presentation	Standard: 200 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only.



G·S·L

WWW.GLOVEMAN.CO.UK

Gloveman Supplies Ltd

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 310734

f. 01209 314760