

G500 Datasheet



AQL 1.5

EN445 Parts 1, 2 & 3

ASTM Method F1671-97

Class 1 Medical Device

MDD 93/42/EEC



Vynite Powder Free Gloves

General Details

Material	Nitrile and Polyvinyl Chloride
Surface Finish	Powder Free
Colour	Pastel Green/Blue
Geometry	Ambidextrous
Texture	Smooth
Cuff and Finishing	Beaded

Glove Specifications

Size	Weight g (M)	Length mm	Palm Width mm	Cuff	Thickness Palm	Finger
XS	5 ± 0.3	230 (min)		0.08	0.08	0.08
S	5.5 ± 0.3	230 (min)	87 ± 3	0.08	0.08	0.08
M	6 ± 0.3	230 (min)	95 ± 3	0.08	0.08	0.08
L	6.5 ± 0.3	230 (min)	106 ± 3	0.08	0.08	0.08
XL	7 ± 0.3	230 (min)	115 ± 3	0.08	0.08	0.08
Tensile strength at break			13MPa (min)			
Elongation at break			300%			

Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

Manufacturing Standards

Maunufacturing Standards	Incompliance with ISO2859
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only. Avoid high polar solvents.



G·S·L
WWW.GLOVEMAN.CO.UK

Gloveman Supplies Ltd
Vision House, Treleigh Ind Est,
Redruth TR16 4AX
t. 01209 310734
f. 01209 314760