

G15 Datasheet



AQL 1.5

EN445 Parts 1, 2 & 3

EN388

EN374

Class 1 Medical Device

MDD 93/42/EEC



General Details

Material	PVC plastisol
Surface Finish	Powder Free
Colour	Clear
Geometry	Ambidextrous
Texture	Smooth
Cuff and Finishing	Beaded

Glove Specifications

Size	Weight g/pc	Length mm	Palm Width mm	Thickness		
				Cuff	Palm	Finger
XS	3.5	230	75 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
S	4	230	85 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
M	4.5	230	95 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
L	5	230	105 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
XL	5.5	230	115 ± 5	0.06 ± 0.02	0.09 ± 0.02	0.11 ± 0.02
Tensile strength before ageing (all gloves)				Load @ peak MPA	Min.13	
Tensile strength after ageing (all gloves)				Load @ peak MPA	Min.13	

Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

Manufacturing Standards

Maunufacturing Standards	Incompliance with ISO 9001:2000
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only. Avoid high polar solvents.



G·S·L

WWW.GLOVEMAN.CO.UK

Gloveman Supplies Ltd

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 310734

f. 01209 314760

Powder Free Vinyl Exam Gloves