

# G1 Datasheet Natural Latex Powder Free Gloves

AQL 1.5

EN455 Parts 1, 2 & 3

FSC Latex

MDD 93/42/EEC

EN374

Class 1 Medical Device



## General Details

Material	Natural Latex
Surface Finish	Powder Free (inner chlorinated., outer polymer coated)
Colour	Natural
Geometry	Ambidextrous
Texture	Fully Textured (Palm Textured)
Cuff and Finishing	Beaded

## Glove Specifications

Size	Weight g/pc	Length mm	Palm Width mm	Thickness		
				Cuff	Palm	Finger
XS	4.8 ± 0.2	240	≤ 80	0.09 ± 0.01	0.13 ± 0.01	0.15 ± 0.01
S	5.3 ± 0.2	240	85 ± 5	0.09 ± 0.01	0.13 ± 0.01	0.15 ± 0.01
M	5.8 ± 0.2	240	95 ± 5	0.09 ± 0.01	0.13 ± 0.01	0.15 ± 0.01
L	6.4 ± 0.2	240	105 ± 5	0.09 ± 0.01	0.13 ± 0.01	0.15 ± 0.01
XL	6.9 ± 0.2	240	≥ 110	0.09 ± 0.01	0.13 ± 0.01	0.15 ± 0.01
Powder Content (Maximum)			2 mg / glove			
Powder Content (Maximum)			50 µg / g			

## Performance Requirements

Test	Related Defects	Inspection Level	AQL
Water Leak at Final Release	Holes	G-1 Single Sampling Plan	1.5

## Manufacturing Standards

Manufacturing Standards	Incompliance with ISO 9001:2008
Presentation	Standard: 100 gloves per dispenser box; 10 dispenser boxes per Carton
Glove Care	Should not store under direct sunlight or at high heat and humid conditions. Single use only.



**G·S·L**

WWW.GLOVEMAN.CO.UK

**Gloveman Supplies Ltd**

Vision House, Treleigh Ind Est,

Redruth TR16 4AX

t. 01209 310734

f. 01209 314760